



Ricardo Morren  
Juvo Solutions,  
PO Box 82,  
Bentley, WA 6982

22<sup>nd</sup> January 2015

Dear Ricardo,

Reference: testing the Juvo Solutions JH6060 Transfer bench to 200 kg user weight, our Laboratory reference is 492394.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS/NZS 3973:2009, AS/NZS 4688.2-2000 and AS/NZS 4688.3-2000.

The official ten (10) page report, dated 22<sup>nd</sup> January 2015 and referencing job number 492394, identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Juvo Solutions JH6060 Transfer bench as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of AS/NZS 3973:2009, AS/NZS 4688.2-2000 and AS/NZS 4688.3-2000 (excluding the methods indicated in the report as "not tested") for user mass 200 kg.

Compliance is conditional on the appropriate informative labelling and marking requirements as detailed in the official report being met with all products made available for sale.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel  
Senior Test Technician  
Authorised Signatory

**NovitaTech (a division of Novita Children's Services)**  
171 Days Road, Regency Park, South Australia 5010  
PO Box 2438, Regency Park, South Australia 5942  
Telephone 1300 855 585 • Fax (08) 8243 8208  
enquiries@novita.org.au • www.novitatech.org.au  
ABN 53 137 537 636

