



Ricardo Morren, Operations Manager
Juvo Solutions
18 Mallaig Way,
Canningvale, WA 6155

11th July 2014

Dear Ricardo,

Reference: review of compliance of the following Juvo Solutions Self-propelled mobile shower commodes with sliding footplates:

- JH3046, 46 cm seat width, 200 kg user weight
- JH3051, 51 cm seat width, 300 kg user weight
- JH3056, 56 cm seat width, 300 kg user weight
- JH3061, 61 cm seat width, 300 kg user weight
- JH3066, 66 cm seat width, 300 kg user weight

This review is based on testing performed on the JH1000 and JH4000 range of mobile shower commodes (up to 66cm seat width, 300 kg user weight), our Laboratory reference is 492257.

This letter is to advise you that this Laboratory has reviewed the above listed Mobile shower commodes to the following Standard method: AS/NZS 3973:2009 Shower / toilet chairs – Mobile and static (except corrosion resistance and surface finishes).

These JH3000 series models use identical frame, wheel and footrest components to the JH1000 and JH4000 series of chairs that were tested by this laboratory (Our references 491546 and 492255 for 66cm width models)

It is the determination of this Laboratory that the JH3046 Self-propelled mobile shower commode (200kg) and JH3051, JH3056, JH3061 and JH3066 (300 kg) comply with the relevant requirements of AS/NZS 3973:2009 Shower / toilet chairs – Mobile and static (excluding the methods indicated in the report as “not tested”) for user masses indicated.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician,
Authorised signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

