



Ricardo Morren,
Operations Manager
Juvo Solutions
18 Mallaig Way,
Canningvale, WA 6155

19th October 2015

Dear Ricardo,

Reference: testing the Juvo Solutions JH2036 Attendant propelled shower chair / commode (Child size, 136 kg user weight), our Laboratory reference is 492502.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS/NZS 3973:2009 Shower / toilet chairs – Mobile and static (except corrosion resistance and surface finishes).

The official ten (10) page report, dated 19th October 2015 and referencing job number 492502 identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Juvo Solutions JH2036 Attendant propelled shower chair / commode (Child size) as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of AS/NZS 3973:2009 Shower / toilet chairs – Mobile and static (excluding the methods indicated in the report as “not tested”) for user mass 136 kg.

Compliance is conditional that all products made available for sale must have permanent marking with the following details:

- Registered name/trademark and address of the manufacturer or distributor
- Recommended maximum user weight (136 kg)
- The chair is suitable for use in showers and bathrooms
- The chair is suitable for wet areas but not pools and similarly corrosive environments

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Wayne Wurfel
Senior Test Technician,
Authorised signatory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

